



OLS-85/OLS-85P

SmartClass™ Fiber Optical Light Sources

Operating manual

BN 2313/98.21

2019.12

English

Please direct all inquiries to your local Viavi sales company. The addresses can be found at:

www.viavisolutions.com/en-us/contact-sales-expert

A description of additional instrument features can be found at:

www.viavisolutions.com/en-us/products/network-test-and-certification

Notice

Every effort was made to ensure that the information in this document was accurate at the time of printing. However, information is subject to change without notice, and Viavi reserves the right to provide an addendum to this document with information not available at the time that this document was created.

General Public License

This product includes Qt 4.8.3 software licensed under the GNU LGPL v. 2.1, and barebox 2012.02.0 and linux 3.1.0-rc8 software licensed under the GNU GPL v. 3. To request a copy of the source code for this software, please write to:

Open Source Request • Viavi Legal Department •
430 N. McCarthy Blvd. • Milpitas, CA 95035 USA

Copyright

© Copyright 2019 Viavi Solutions Inc. All rights reserved. Viavi and the Viavi logo are trademarks of Viavi Solutions Inc.

All other trademarks and registered trademarks are the properties of their respective owners.

Viavi Solutions Deutschland GmbH
Arbachtalstraße 5, D-72800 Eningen u. A.

Order no.: BN 2313/98.21

Version: 2019.12

Previous version: 2019.06

Notes:

Changes may be made to specifications, descriptions and delivery information.

CONTENTS

INTRODUCTION	5
OLS-85/OLS-85P Light Sources	5
Operating manual update	7
Symbols used in this operating manual	8
SAFETY INFORMATION	10
Warning symbols on the unit	10
Proper usage	10
Laser safety	11
Battery operation	11
Ventilation	12
PS4 Universal AC/DC Power Supply	12
GETTING STARTED	14
Unpacking the instrument	14
Connecting optical cables	22
BASIC OPERATION	24
Switching the instrument on/off	24
Operator control panel	24
Menus and display elements	25
Navigating in the menus	28
Displaying application information	28
Changing System Settings	29
WiFi menu (optional)	33
Bluetooth menu (optional)	33
Installing a software option	33
Updating the firmware	34
Creating screenshots	34
MANAGING PROJECTS	36
Creating a new project	36
Editing a project	37
Deleting a project	38
Activating a project	38
SOURCE OPERATION	39
General information	39
Selecting an output port and wavelength	41
About Auto-I mode	41
Selecting a modulation or Auto-I mode	42
Setting the power level	43

Switching on/off the laser.....	43
PROBE/PCM OPERATION	45
General information.....	45
Patch Cord Microscope (PCM).....	45
FMAE series adapters for the PCM	46
Features available with the P5000i Digital Probe	46
QuickCapture™ key	48
File toolbar.....	49
Configuring the Digital Probe	49
Saving Probe/PCM results	53
DATA MANAGEMENT.....	55
Saving results	55
Data management of Probe and PCM tests	55
Exporting results to USB	58
Making a report	59
MAINTENANCE	60
Cleaning the test port	60
Cleaning the instrument	60
REMOTE CONTROL.....	61
ENVIRONMENTAL COMPLIANCE.....	62
SPECIFICATIONS.....	64
OLS-85 specifications	64
General specifications	66
ORDERING INFORMATION	67
OLS-85/OLS-85P stand alone units.....	67
Included items	67
Accessories	67

1 INTRODUCTION

OLS-85/OLS-85P Light Sources

The SmartClass™ Fiber OLS-85 light sources are professional, versatile, compact handheld instruments designed for qualification and certification of fiber optic networks. Thoroughly chosen combinations of available wavelengths make the SmartClass™ Fiber OLS-85 light sources the optimum choice for link loss testing and characterization of long-haul, metro, and access telecommunication networks, as well as for data center and local area network testing.

Single-Mode (SM) and Multi-Mode (MM) – The all-in-one solution

OLS-85 devices that have both single mode and multi-mode light sources are the most flexible test instruments. These instruments are particularly suitable for rental and measurement service companies.

Auto λ – Do it right the first time

OLS-85 Auto λ feature sends information about the emitted wavelengths to a receiving-end (e.g. OLP-85), enabling the receiving-end to automatically set up for correct wavelength setting and hence ensuring a correct measurement the first time.

Multi λ – Reduce testing time to a minimum

Combined with a compatible power meter (e.g. OLP-85) Multi- λ testing reduces testing time to a minimum by simultaneously testing at all wavelengths.

Macrobending test ready

Several OLS-85 variants provide a 1625 nm light source which is the ideal wavelength for detecting macrobending issues in fiber links.

Fiber inspection ready – Inspect before you connect

The SmartClass™ Fiber OLS-85 instruments are prepared for fiber inspection. Connection of a P5000i Digital Probe to the OLS-85 enables a technician to perform best practices fiber

inspection and automated Pass/Fail testing of optical connectors/adapters in order to ensure industry standard fiber endface quality and cleanliness.

You don't need an additional fiber microscope set – simply plug in a P5000i Digital Probe.

Report generation – Do it the easy way

Of course, it is possible to make measurement reports with a PC tool.

But it's even easier with SmartClass™ Fiber – generate your report on the instrument directly and you're done.

FTTx/PON aware

OLS-85 instruments allow for passive optical network (PON) testing at the ITU-T G.983.3, and IEEE 802.3ah specified wavelengths 1310 nm, 1490 nm, and 1550 nm. This is the ideal source for PON network loss testing.

Rugged and portable

Battery operation from AA dry batteries, or from a rechargeable Li-Ion Battery Pack ensure a long operating time in the field and a robust, shock-proof design makes the SmartClass™ Fiber instruments the perfect choice for optical network field testing, even under tough conditions.

Remote controllable

Operation from an AC/DC power supply and remote control capabilities via USB 2.0 and Ethernet make the SmartClass™ Fiber instruments even a perfect solution for fixed installations in central offices, in production environments, and on the laboratory workbench.

Differences between the instruments

	BN 2313/xy				
	01	05	06	22	26
Wavelengths					
850 nm	–	✓	–	–	–
1300 nm	–	✓	–	–	–
1310 nm	✓	✓	✓	✓	✓
1490 nm	–	–	✓	–	✓
1550 nm	✓	✓	✓	✓	✓
1625 nm	–	–	✓	✓	✓
Optical outputs					
Single-mode port	PC ¹⁾	PC	PC	APC ²⁾	APC
Multi-mode port	–	PC	–	–	–

1) physical contact

2) angled physical contact

Operating manual update

Continuing enhancement and further development of the SmartClass family may mean that this operating manual does not cover all the latest functions of your instrument.

If the operating instructions about features supported by your instrument are missing, please visit the Viavi web site to check if additional information is available.

To download the latest user manual:

1. Visit the Viavi web site at <http://updatemyunit.net>.
2. Select your SmartClass™ Fiber model from the product line.
3. Open the download area and download the latest user manual.

Symbols used in this operating manual

Various elements are used in this operating manual to draw attention to special meanings or important points in the text.

Symbols and terms used in warnings

The following warnings, symbols and terms are used in this document in compliance with the American National Standard ANSI Z535.6-2011:

NOTICE

Follow the instructions carefully to avoid **damage to or destruction of the instrument**.

CAUTION

Follow the instructions carefully to avoid a low or medium risk of **injury to persons**.

WARNING

Follow the instructions carefully to avoid **severe injury** to persons.

DANGER

Follow the instructions carefully to avoid **death or severe injury** to persons.



High voltage

Follow the instructions carefully to avoid **damage** to the instrument or **severe injury** to persons.

This safety instruction is given if the danger is due to **high voltage**.



Laser

Follow the instructions carefully to avoid **damage** to the instrument or **severe injury** to persons.

This safety instruction is given if the danger is due to **laser radiation**. Information specifying the laser class is also given.

Warning format

All warnings have the following format:


WARNING

Type and source of danger

Consequences of ignoring the warning.

- Action needed to avoid danger.

The following elements are used in this operating manual:

✓	Requirement This requirement must be met first; e.g. ✓ The system is switched on.
▶ 1. 2.	Instruction Follow the instructions given (the numbers indicate the order in which the instructions should be followed); e.g. ▶ Select mode.
<i>Italics</i>	Result Indicates the result of following an instruction; e.g. <i>The page opens.</i>
Boldface	Pages, controls, and display elements Screen pages, controls, and display elements are indicated in boldface .
Text in blue	Cross references Cross references are indicated in blue type. When using the PDF version, just click on the blue text to skip to the cross reference.
[]	Instrument keys Instrument keys are indicated within square brackets.
[More]	Touchscreen buttons Touchscreen buttons are indicated within blue square brackets.

2 SAFETY INFORMATION

Warning symbols on the unit



Warning symbols indicating a potential hazard

- ▶ In all cases where the unit is labeled with a warning symbol, the operating manual must be consulted to learn more about the nature of the potential hazard and any action that must be taken.

Proper usage

This instrument is intended for measurements on optical fiber devices and systems.

- ▶ Please make sure the instrument is not operated outside the permitted ambient conditions.
- ▶ Always make sure that the instrument is in proper working order before switching it on.

Laser safety



⚠ WARNING

Dangerous laser radiation

Laser radiation can cause irreparable damage to eyes and skin.

- ▶ This device is a Class 1 laser product according to DIN EN 60825-1:2007.
 - ▶ Always be aware of the hazard level of the instrument to be connected.
 - ▶ Connect all optical fibers before switching on the radiation source.
 - ▶ Switch off the laser source before disconnecting the optical fibers.
 - ▶ Never look directly into the unconnected port of the instrument, the output of a laser source, or an optical fiber connected to a source or a system.
 - ▶ Always cover unused ports.
 - ▶ Be aware that the emitted light is not visible.
 - ▶ Heed the normal precautions for working with laser radiation and consider local regulations.
-

Battery operation

⚠ WARNING

Explosion danger

Short-circuiting the batteries can result in overheating, explosion, or ignition of the batteries and their surroundings.

- ▶ Never short-circuit the battery contacts by touching both contacts simultaneously with an electrical conducting object.
 - ▶ Only use AA size dry batteries or rechargeable batteries.
 - ▶ Make sure the batteries are inserted with the correct polarity.
-

Ventilation

NOTICE

Insufficient ventilation

Insufficient ventilation can damage the instrument or adversely affect its function and safety.

- Ensure adequate ventilation when operating the instrument.
-

PS4 Universal AC/DC Power Supply

Safety class

The PS4 Universal AC/DC Power Supply unit has a protective isolation that conforms with IEC 60950.

Environmental conditions

NOTICE

Ambient temperature too high/low

Temperatures outside the operating range of 0 to +40 °C can damage the PS4 Universal AC/DC Power Supply or adversely affect its function and safety.

- Only operate the PS4 Universal AC/DC Power Supply indoors.
The PS4 Universal AC/DC Power Supply must only be operated at ambient temperatures between 0 and +40 °C.

NOTICE

Insufficient ventilation

Insufficient ventilation can damage the PS4 Universal AC/DC Power Supply or adversely affect its function and safety.

- Ensure adequate ventilation when operating the PS4 Universal AC/DC Power Supply.

NOTICE**Condensation**

Operation in the presence of condensation can damage the PS4 Universal AC/DC Power Supply or adversely affect its function and safety.

- ▶ Do not operate the PS4 Universal AC/DC Power Supply if condensation has formed.
 - ▶ If condensation cannot be avoided, such as when the PS4 Universal AC/DC Power Supply is cold and is moved to a warm room, wait until the PS4 Universal AC/DC Power Supply Unit is dry before plugging it into the AC power line.
-

3 GETTING STARTED

Unpacking the instrument

Packing material

We suggest that you keep the original packing material. It is designed for reuse (unless it is damaged during shipping). Using the original packing material ensures that the instrument is properly protected during shipping.

Checking the package contents

- Unpack the instrument and check the package contents. For more information see [“Included items” on page 67](#).

Checking for shipping damage

After you unpack the instrument, check to see if it was damaged during shipping. This is particularly likely if the packaging is visibly damaged. If there is damage, do not attempt to operate the instrument. Doing so can cause further damage. In case of damage, please contact your local Viavi sales company. Addresses can be found at www.viavisolutions.com.

Recovery following storage/shipping

Condensation can occur if a instrument that is stored or shipped at a low temperature is brought into a warm room. To prevent damage, wait until no more condensation is visible on the surface of the instrument before powering it up. Do not operate the instrument until it has reached its specified temperature range and wait until it has cooled down if the instrument was stored at a high temperature (see [“General specifications” on page 66](#)).

Instrument overview



Fig. 1 Front view OLS-85

1	Connector interface
2	Test head cover (green for APC- and gray for PC connectors)
3	3.5 inch touchscreen
4	Key pad (operator control panel)
5	Battery compartment and stand (on rear of instrument)
6	USB 2.0 device port (Type Micro-B)
7	USB 2.0 host port (Type A) and external power supply connector
8	Ethernet port (RJ-45)

Connector panel




Fig. 2 Connector panel of a single port instrument

1 Optical connector




Fig. 3 Connector panel of a dual port instrument

1 Optical connector 1 for single mode

2 Optical connector 2 for multi mode




Fig. 4 External power supply connector and communication interfaces

- | | |
|---|------------------------------------|
| 1 | Ethernet port (RJ-45) |
| 2 | External power supply connector |
| 3 | USB 2.0 host port (Type A) |
| 4 | USB 2.0 device port (Type Micro-B) |

Power supply

The following power sources can be used to operate the OLS-85:

- Eight 1.5 V dry batteries (Mignon AA size, alkaline type recommended)
- Eight 1.2 V NiMH rechargeable batteries (Mignon AA size, no internal charge)
- PS4 Universal AC/DC Power Supply (optional)
- RBP2 Li-Ion Battery Pack (optional)

NOTE: Best Measurement accuracy is achieved when the measurements are performed in battery mode (w/o AC adapter).

Battery operation

WARNING

Dangers when handling batteries

Handling batteries may be dangerous. Please note the following safety instructions.

- ▶ Please note the battery operation safety information in the chapter ["Battery operation" on page 11](#).

Replacing batteries

- ▶ Do not replace individual batteries. Always change all eight batteries at the same time.
- ▶ Always use eight batteries of the same type; i.e. do not mix rechargeable and non-rechargeable batteries.

Replacing batteries



Fig. 5 Replacing the batteries

- | | |
|---|--------------------------|
| 1 | RBP2 Li-Ion Battery Pack |
| 2 | Latch lock |
| 3 | AA battery tray |

The battery compartment is on the back of the instrument.

1. Press down the latch to release and to open the lid of the battery compartment.
2. Insert new batteries in the tray or remove the used batteries from the tray and replace all eight with fresh ones.

NOTE: Take care to insert the batteries correctly. The correct polarity is indicated by a diagram inside the battery compartment.

– or –

Insert new or replace the RBP2 Li-Ion Battery Pack.

– or –

Switch from non-rechargeable batteries to rechargeable ones by replacing the battery tray with a new battery pack (or vice versa).

3. Close the battery compartment.

4. Press the **[ⓘ]** key to switch on.

Recharging the batteries

The rechargeable RBP2 Li-Ion Battery Pack is recharged when the PS4 Universal AC/DC Power Supply is being used to power the instrument. The instrument switches to trickle charging automatically as soon as the RBP2 Li-Ion Battery Pack is fully charged.

NOTE: Rechargeable AA batteries will not be recharged in the instrument. For AA-type rechargeable batteries please use an external charger.
It is not possible to charge the rechargeable AA batteries or the RBP2 Li-Ion Battery Pack via the USB interface.

The instrument is powered by the PS4 Universal AC/DC Power Supply if the PS4 Universal AC/DC Power Supply and the USB interface are both connected.

Deep discharge

A rechargeable battery that appears to be dead (unit will not turn on even when connected to external power) may well be in a deep discharge state. It can be fully charged with the following charging cycle:

1. Plug in the power cord for 1–1.5 hours.

Battery should be partly charged now.

2. Switch the power on.

If the instrument starts: Let the battery fully charge.

– or –

If the instrument does not start:

Unplug power cord, plug it back in and repeat steps 1 and 2.

NOTE: If the battery is completely deep-discharged, this cycle might need to be repeated up to 3 times.

General tips on using batteries

- Always handle batteries with care.
- Do not drop or damage the batteries or expose them to excessively high temperatures.
- Do not store the batteries for more than one or two days at very high temperatures (e.g. in a vehicle), either separately or inserted in the instrument.
- Do not leave discharged batteries in the instrument for a long time if it is not being used.

Other basic safety precautions are as follows:

- Do not use PS4 Universal AC/DC Power Supply outdoors or in wet or damp locations.
- Connect the PS4 Universal AC/DC Power Supply to the correct mains voltage, as indicated on the rating label.
- Do not allow anything to rest on the power cord, and do not place the product where people can walk on the power cord.
- Avoid using this product during electrical storms. There may be a remote risk of electric shock from lightning.
- Do not use this product in the vicinity of a gas leak or in any explosive environment.
- Do not attempt to service this product yourself, as opening or removing covers may expose you to dangerous, high voltage points, and other hazards. Contact qualified service personnel for all service.

Environmental protection

Please dispose of any unwanted dry batteries and rechargeable batteries carefully. They should also be removed from the instrument if it is to be discarded. If facilities in your country exist for collecting such waste or for recycling, please make use of these rather than throwing the batteries in with normal trash. You will often be able to return used batteries to the place where you purchase new ones. Any dry or rechargeable batteries that you purchased from Viavi can be returned to one of our Service Centers for disposal.

Operation from AC power

NOTE: Only the PS4 Universal AC/DC Power Supply may be used to operate the OLS-85 from AC power.

To fit the AC line plug adapter:

1. Select the appropriate AC line plug adapter.

- Slide the AC line plug adapter into the slot.
The PS4 Universal AC/DC Power Supply is ready for use.




Fig. 6 PS4 Universal AC/DC Power Supply

To change the AC line plug adapter:

- Squeeze both sides of the PS4 latch lock (see Fig. 6).
- Push the AC line plug adapter upwards.
- Slide a different AC line plug adapter into the slot (see Fig. 7).




Fig. 7 PS4: Changing the AC line plug adapter

To operate the OLS-85 from AC power:

- Connect the PS4 DC power cord to the OLS-85 external power supply connector.
(The connector is under the cover on the right side.)
- Plug the PS4 into the AC line socket.

The OLS-85 switches on automatically when powered from the PS4.

NOTE: The PS4 provides power even if dry or rechargeable batteries are inserted in the instrument.
The OLS-85 cannot be powered via the USB interface.

Connecting optical cables

Mounting test adapters

Viavi provides a number of test adapters for connecting the OLS-85 to the interface to be tested.

You can connect all standard optical connector types to the instrument using these adapters. The test adapters are suitable for connectors with planar (PC) and angled end surfaces (APC). Contact your local Viavi sales company for available adapter types.

The OLS-85 connector type must match the cable connector type:

- PC: BN 2305/01/11
- APC: BN 2305/21/26/31/36/66

The PC/APC versions are easily identified by the colors of the name labels on the front:

- PC = blue
- APC = green

NOTE: Only single mode fibers (SMF) may be connected to the OLS-85. To mount the JAE test adapter:

- ✓ The optical connectors are properly cleaned (see [“Cleaning the test port” on page 60](#)).
- 1. Open the head cover and remove the protective cap (if still mounted).
- 2. Place the test adapter vertically on the optical connector with the locking lever open.
- 3. Close the locking lever once the test adapter is firmly seated. You will hear the locking mechanism click.
- 4. Repeat the procedure if the instrument is equipped with two ports.
- 5. Fit the fiber optic cable to the test adapter or close the head cover.




Fig. 8 Mounting the JAE test adapter

To mount the SENKO test adapter:

1. Open the head cover and remove the protective cap (if still mounted).
2. Unscrew the SENKO test adapter and pull it off vertically.
3. Place the SENKO test adapter vertically on the optical connector.
4. Fix the SENKO test adapter with two screws.
5. Repeat the procedure if the instrument is equipped with two ports.
6. Fit the fiber optic cable to the test adapter or close the head cover.



Fig. 9 Mounting the SENKO test adapter

4 BASIC OPERATION

Switching the instrument on/off











To switch the instrument on:

- ▶ Press the [ⓘ] key to switch on the instrument.

To switch the instrument off:

- ▶ Press the [ⓘ] key to shift the instrument into hibernate mode.
 - or –
- 1. Hold the [ⓘ] key to open the power off menu.
- 2. Tap the [Power Off] button or press the central key to switch off the instrument.

Operator control panel

	Press to go to the homescreen.
	Press to open menu.
	Press to go back within an application or cancel input.
	Press to toggle between the inspect view and the test view.
	Press to switch the instrument on and off. LED glows green when the instrument is on.
	Press an arrow key to: <ul style="list-style-type: none"> • navigate through the menus • change values in the menus Press the central key to: <ul style="list-style-type: none"> • confirm the selection
	Press to save results.
	Glows red when battery is low.
	Glows red when a measurement is running in the background.
	Glows orange when battery is charged; flashes orange when battery is charging. Turned off when dry batteries are used or battery bay is empty.

Menus and display elements

Home screen




Fig. 10 Home screens: Test applications (screen shows PCM version) (1), Management (2), Settings (3)

	Menu	See page
	Test applications In this menu you can open test applications.	26
	Management In this menu you can: <ul style="list-style-type: none"> • manage projects • define a test profile (not supported yet) • get access to stored data 	36 and 55
	Settings In this menu you can: <ul style="list-style-type: none"> • manage WIFI and Bluetooth settings • change instrument settings • open the help • display instrument information • manage software options 	29

Test applications


















Fig. 11 Test applications (screen shows PCM version)

	Application	See page
	Source To edit the laser for maintenance and characterization of single-mode and multi-mode networks.	39
	Probe To view and inspect the bulkhead (female) connectors.	45
	PCM (Patch Cord Microscope) (only BN 2326) To view and inspect the patch cord connector of the fiber.	45

Elements in the top bar

	Project title Indicates the title of the active project
	WIFI Indicates that Wifi is installed. It does not indicate an active Wifi connection.
	Bluetooth Indicates that Bluetooth is installed. It does not indicate an active bluetooth connection.
	External power supply The MPOLx-85 is powered by the external AC adapter when this symbol is shown.
	Battery status Indicates the battery charge status. If it is not shown, only the AC adapter is active. <div>  Charge state = 100% </div> <div>  Charge state = 25% </div> <div>  Battery is charging. </div>
	Auto-Off Indicates whether the instrument turns off within a certain time.
 	Active laser source Blinking yellow-red indicates that the built-in laser source is active.

Navigating in the menus

- ▶ Press the **[↵]** key to open the home screen.
Depending on the active project type the Projects tab shows the available test applications or the active Workflow project.
- ▶ Press the **[☰]** key to open the context-sensitive menu.
Depending on which application is selected, a different menu opens.

To select a menu item:

1. Press the arrow keys to highlight an item.
2. To confirm, press the center key within the arrow keys.
 - or –
 - Tap the desired button on the touchscreen.

To leave a menu without making any changes:

- ▶ Press the **[↶]** key.

NOTE: All actions can be operated via the operator control panel or the touchscreen. The following instructions describe only touchscreen operation.

To toggle between different display modes:

- ▶ Tap the display to toggle between the display modes (context-sensitive).

Displaying application information

The Info menu provides information about the displayed test application.

✓ The instrument is in the chosen application mode.


1. Press the **[☰]** key and tap the **[More]** button.
2. Tap the **[Info]** button.
The application information is displayed.

Changing System Settings


In the **Settings** menu you can change instrument settings, get information and help about the instrument, or update the firmware.

To open the Settings menu:

✓ The home screen is displayed.


1. Select the  tab.

*The **Settings** menu opens:*









2. In the  menu tap the **[More]** button.


More setting options are displayed.



The following table gives a short overview of the menu items. These are explained in the following sections.

Icon	Function	See page
	WiFi To configure the wireless local area network. The last selected item is displayed in magenta.	33
	Bluetooth To configure the Bluetooth interface.	33
	Brightness To adjust the display brightness.	30
	Help To show device information	30
	SW Options To check and install software options.	Not yet available, 33
	More To access further instrument settings. See following sections.	30

Adjusting the display brightness

1. In the  menu tap the **[Brightness]** button.
2. Tap a symbol to change the display brightness.
3. Tap **[OK]**.


Showing device information

- In the  menu tap the **[Help]** button.
Information about the hardware and software is displayed.

Setting the **[Auto-Off]** interval

When **[Auto-Off]** is set, the device will switch off after the selected interval without any user action.


NOTE: **[Auto-Off]** is only active when no external power supply is connected.

1. In the  menu tap the **[More]** button, then tap **[Auto-Off]**.
2. Select the desired interval.

Setting the [Screen-Off] interval

When [Screen-Off] is set, the display will switch off after the selected interval without any user action.


NOTE: [Screen-Off] is only active when no external power supply is connected.

1. In the  menu tap the [More] button, then tap [Screen-Off].
2. Select the desired interval.


To switch on the display after it was switched off:

- ▶ Just tap the display to switch it on.

Selecting a language

1. In the  menu tap the [More] button, then tap [Language].
2. Tap the desired language.
After selecting a new language the instrument has to be restarted.
3. Tap [Yes] to restart the instrument.
– or –
Tap [No] to keep the previous language.

Setting date & time

- ▶ In the  menu tap the [More] button, then tap [Date & Time].

To set the date:

1. Tap [Date].
2. Tap [Day], [Month], or [Year].
3. Type in the desired value.
4. Tap [OK].

To set the date format:

1. Tap [Date Format].
2. Select the desired date format.


To set the time:

1. Tap [Time].
2. Tap [Hour], [Minute], or [Second].
3. Type in the desired value.
4. Tap [OK].

To set 24-hour or 12-hour time:

1. Tap [Time Format].
2. Select the desired time format.

Setting the Ethernet protocol

- In the  menu tap the [More] button, then tap [Ethernet].

To select the IP mode:

1. Tap the [IP Mode] button.
2. Select the desired Ethernet mode.

To set the IP address:

- ✓ The IP mode **Static** is selected.

 1. Tap [IP Address].
 2. Type in the desired IP address.
 3. Tap [OK].

To set the gateway:

- ✓ The IP mode **Static** is selected.

 1. Tap [Gateway].
 2. Type in the desired gateway.
 3. Tap [OK].


To set the netmask:

- ✓ The IP mode **Static** is selected.

 1. Tap [Netmask].
 2. Type in the desired netmask.
 3. Tap [OK].


Calibrating the touchscreen

NOTE: Using a pen when calibrating the touchscreen will improve the calibration accuracy.


1. In the  menu tap the [More] button, then tap [Touchscreen calibration].
2. Follow the given instructions.

Resetting to the factory default values


NOTE: Setting the factory default values does not affect your stored measurement results.

1. In the  menu tap the [More] button, then tap [Factory Reset].
2. Tap [Yes] to proceed.
 - or –
 - Tap [No] to cancel.



WiFi menu (optional)

- ▶ In the  menu tap the [WiFi] button.
Editable settings are displayed
 - Enabled: switch WiFi on/off
 - Mode: switch between Ad-Hoc / Access PointTap [Enabled] to switch WiFi on.
Information on enabled wifi connection is displayed
 - SSID:
 - IP Address
 - Netmask
 - MAC Address
 - Port

Bluetooth menu (optional)

- ▶ In the  menu tap the [Bluetooth] button.
editable settings are displayed
 - Enabled: switch Bluetooth on/offTap [Enabled] to switch Bluetooth on.
Information on enabled bluetooth connection is displayed
 - SSID
 - MAC Address
 - PIN Code for connection establishment

Installing a software option

1. Connect an USB stick with the corresponding option file.
2. In the  menu tap the [Software Options] button.
3. Press the  key.
4. Tap [Install from USB].
The license key is read in from the root directory of the USB stick.
The option(s) are installed.




Updating the firmware

The latest version of the firmware can be downloaded from the Internet.

To find the latest firmware version:

1. Visit the Viavi web site at <http://updatemyunit.net>.
2. Select your model from the product line.
3. Open the download area and download the latest firmware.
The downloaded update.exe file needs to be executed on a PC.
4. Unzip and save the individual update files to the root directory of an USB flash drive.




To start the firmware update:

- ✓ The update files have been extracted and are located on the USB flash drive.
 - ✓ The instrument is fully charged or the PS4 Universal AC/DC Power Supply is connected.
 - ✓ The instrument is switched off and not in hibernate mode.
1. Plug the USB flash drive with the extracted update files into one of the instrument USB ports.
 2. Hold down the  key and shortly press the  key to switch the instrument on.
The firmware update screen appears.
 3. Release the  key.
 4. Press the center key within the arrow keys to start the update procedure.
 5. After Startup, when the Homescreen is displayed, remove the USB flash drive.

Creating screenshots

Screenshots can be stored directly on an USB stick.

To create a screenshot:

1. Connect a USB stick to one of the USB ports on the right side of the device.
The  symbol is displayed at the top of the screen.
2. Long-press both the  and the  keys.
The Save New Screenshot screen is displayed.
3. Enter a file name and use the pencil button to edit the field.

4. Set the **Add Auto-Increment Number** function **ON** or **OFF**.
If the setting is **ON**, a number is added to the proposed name, ascending each time that a new screenshot is taken.
Edit the **Auto-Increment Number** field by tapping the pencil button if you want to change the current number.
5. If you are using both USB ports, select **USB1** or **USB2** by tapping the corresponding button.
6. Tap the **[Save]** button to store the screenshot on the USB stick.
A directory named
Screenshots_<device_type> <serial_number> is created and
the screenshot is saved in PNG format.
Tap the **[Cancel]** button to interrupt the creation of the
screenshot.

5 MANAGING PROJECTS

Stored Measurements are assigned to a project. Therefore a project first has to be created and set to active. As factory setting one Test-Tool project named **default** is available, which can be edited but not deleted. New projects can be created by copying and changing an existing project or by creating a new one.

With the project some basic information like the project name, a description of the project and how the label is build can be saved.





Fig. 12 Project assignments

Creating a new project

1. Press the **[H]** key, select  and tap **[Projects]**.
*The Projects list opens. As factory setting only the Test-Tool project **default** is available.*





Fig. 13 Projects list with default project.

Adding a new project:

2. Press the [≡] key.
The edit menu opens.



3. Tap [Add].
The title edit menu opens.
4. Type in the project title and tap ✓.
The project is created and displayed in the list.

Copying an existing project:

1. Select the project you wish to copy.
2. Press the [≡] key and tap [Copy].
The title edit menu opens.
3. Type in the project title and tap ✓.
The project is created and displayed in the list.


Editing a project

1. Double click the project.
– or –
Select a project using the up/down arrow key and press the central key.
The edit menu opens:


Name	Project name
Description	Project description Will be shown in the projects list below the project name.
Label Prefix	Prefix that will be added to the label.
Label Index	Start index for the label
Auto Increment Label Index	When enabled label index will be incremented automatically with each storage.

2. Edit the fields.

The changes are immediately effective.



3. Press the  button to close the edit menu.

Deleting a project

1. Select the project you wish to delete.
2. Press the  key and tap **[Remove]**.
3. Tap **[Yes]** to permanently delete the project.
The project is deleted and removed from the list.

Activating a project

To assign measurements to a project, the project must be activated. There can always be only one active project.

1. Select the project you wish to activate.
2. Tap the  button.
– or –
Press the  key and tap **[Set Active]**.
*Allow a few seconds until the label **ACTIVE** is displayed at the selected project.*

NOTE: Savings of measurement results are assigned to the selected project.

6 SOURCE OPERATION

General information

Various fiber optic tests (e.g. loss test, fiber identification) require a reliable and stable optical source. OLS-85/OLS-85P devices are high-quality optical sources, which in addition to a stable optical signal have other useful properties, which simplify and shorten the work processes and exclude potential sources of error in measurement setups.

OLS-85/OLS-85P devices include several light sources – i.e. lasers and/or LEDs – which are stabilized using an automatic gain control mechanism. LEDs are typically used for multi-mode fiber measurements. On the other hand, lasers are used for single-mode fiber measurements.

NOTE: In this manual, the word laser is used synonymously for laser and LED sources.

The output signals of the light sources may be either continuous wave (CW) or modulated at certain frequencies, i.e. 270 Hz, 1 kHz, or 2 kHz. Viavi power meters (e.g. OLP-3x, OLP-8x, MTS-2k/4k/6k) are capable of detecting these modulation frequencies. This method improves ambient light rejection in certain measurement setups. Additionally, the modulated signals can be used for fiber identification or for confirmation of continuity.

OLS-85/OLS-85P instruments provide the Auto λ feature. When activated, the OLS-85/OLS-85P embeds wavelength ID information into the emitted signal(s). This wavelength ID information can be utilized by a compatible Viavi power meter (e.g. OLP-85) for automatic wavelength setup and for simultaneous reception of multiple wavelengths. This minimizes test time and prevents common errors such as incorrect wavelength setting on the optical power meter.

Furthermore, all OLS-85/OLS-85P instruments provide Auto- λ -Multi and Auto- λ -Serial modes, which significantly reduce test time compared to manually switching wavelengths.

Display elements




Fig. 14 Source operation display, laser off (left) and on.

1		Output port buttons Port not selected / port selected / port selected and laser on J Tap to select/deselect.
2	 	Mode button (modulation) Constant wave > 270 Hz > 1 kHz > 2 kHz > ... J Tab button to skip through options. Mode button (Auto-λ) Auto-λ (with one wavelength) > Multi-λ > Serial-λ > Auto-λ J Tap button to skip through options.
3		Wavelength buttons Wavelength selected / wavelength selected and laser on J Tap wavelength to select/deselect.
4		Laser button Laser off / on J Tap button to switch laser on/off.

Selecting an output port and wavelength

1. To select output, tap SM or MM on top of the display.
– or –
Press the [≡] key, tap [Output Selection] and tap the desired output.
The menu is closed and the measurement display is shown.
2. To select a wavelength, tap the desired wavelength(s) on the measurement display.
– or –
Press the [≡] key, tap [Source Selection] and tap the desired wavelength(s).
If Auto-λ is not selected, only one wavelength can be selected. To select Auto-λ see "About Auto-λ mode".
3. Press the [↶] key.
4. To turn on the selected lasers, tap the [Laser On/Off] button or press the central key.

About Auto-λ mode

When in one of the Auto-λ modes, the OLS-85/OLS-85P embeds wavelength ID information into the emitted signal(s). This wavelength ID information can be utilized by a compatible Viavi power meter (e.g. OLP-85) for automatic wavelength setup and for simultaneous reception of multiple wavelengths.

The OLS-85/OLS-85P provides two Auto-λ modes:

- **Multi-λ:** Multi-λ is a proprietary Viavi solution that reduces testing time to a minimum by simultaneously testing at both wavelengths. When activated both wavelength are measured and displayed simultaneously.
Auto-λ-Multi should be selected every time the optical powermeter supports the Auto-λ-Multi feature.
- **Serial-λ** automatically toggles between the selected wavelengths. Combined with a compatible power meter (e.g. OLP-3x, OLP-85), the power meter automatically sets up with correct wavelength settings.
Serial-λ should be selected when low levels are expected on the optical power meter or when the power meter doesn't support Multi-λ.

Wavelengths cannot be reliably detected in following situations:

- The receive level is too low.
- The wavelength encoding cannot be detected due to interference.
- You are measuring the absolute level of a system that does not have wavelength encoding matching Viavi power sources.

Selecting a modulation or Auto- λ mode

Modulation or Auto- λ mode can be selected by the mode button or via the menu. Both ways are described

Enabling modulation and selecting a frequency

Using the [Mode selection] button:

NOTE: A modulation frequency can only be selected when just one wavelength is selected.

1. When two wavelengths are selected first tap one of the wavelengths to deselect it.
2. Tap the [Mode selection] button to toggle between the modulation frequencies (and the Auto- λ mode):




Using the menu:

- Press the [≡] key, tap the [Modulation] button and tap the desired frequency.

The measurement display is shown and the selected frequency is enabled. If Auto- λ was active before, the mode is set to modulation now.

Enabling Auto- λ

Using the [Mode selection] button:

- ✓ A modulation frequency is selected.
- Tap the [Mode selection] button until Auto- λ  is displayed. *Auto- λ is enabled now for the selected wavelengths.*


Using the menu:

- ✓ A modulation frequency is selected.
- Press the [≡] key, tap the [Auto λ] button and tap [Serial- λ].

Enabling Multi- λ or Serial- λ (two wavelengths)




NOTE: Multi- λ and Serial- λ can only be enabled when two wavelengths are selected.

Using the [Mode selection] button:

1. Tap the [Mode selection] button until Auto- λ  is displayed.
2. Activate more than one wavelength by tapping.
3. Tap the [Mode selection] button to select Multi- λ or Serial- λ :






Using the menu:

1. Press the  key, tap the [Auto λ] button and select the desired Auto- λ mode.
The measurement display is shown, but still Auto- λ is enabled.
2. Press the  key, tap the [Source Selection] button and switch on both wavelengths.
3. Press the  key.
The measurement display is shown and the selected Auto- λ mode is enabled.

Setting the power level

The power level can be set for each wavelength separately and will be stored permanently.

Changing the power level


1. Press the  key and tap the [Level Adjustment] button.
2. Select SM or MM for single- or multi-mode and tap the level adjustment button of the desired wavelength.
The display changes to edit mode.
3. Press the arrow keys to increase or decrease the power level:
 - left/right arrow keys: change level in 1 dBm steps
 - up/down arrow keys: change level in 0.01 dBm steps.
4. Press the  key to leave edit mode of the selected wavelength.
5. Press the  key again to leave the edit menu.

Switching on/off the laser

- To switch on the laser, tap the [Laser On/Off] button.
 - or –

Press the center key within the arrow keys.

– or –

Press the  key and tap the **Output On** button.

- To switch off the laser, repeat one of the actions mentioned above.

7

PROBE/PCM OPERATION

General information

Dirty and/or damaged connectors are often the root cause of optical network problems. The Probe and PCM applications enable industry standard inspection and automated Pass/Fail testing with report generation of optical connectors/adapters in order to ensure industry standard fiber endface quality and cleanliness.

For best workflow efficiency, there are two variants of fiber microscopes available in the SmartClass™ Fiber family. The integrated Patch Cord Microscope (PCM) which is best suited for inspecting fiber endfaces of patch cords and the external P5000i Digital Probe (see ["Digital Probe" on page 67](#)) which may be either used for bulkhead inspection or patch cord inspection.

The shortest inspection time is achieved when using one of the OLP-87P models with an integrated PCM for patch cord inspection and a P5000i Digital Probe for bulkhead inspection. Both applications – Probe and PCM – behave essentially the same and are described together below.

NOTE: Only if a P5000i Digital Probe is connected to the instrument is the Probe application fully functional.
Select an OLP-87P version in order to get PCM functionality.

Patch Cord Microscope (PCM)

The PCM is a microscope used to view and inspect patch cord (male) sides of fiber connectors.

In order to support a wide variety of fiber optic connectors the PCM provides an exchangeable FMAE adapter. The dedicated QuickCapture key provides either instant triggering of a Pass/Fail test or freezing the live image. For best workflow adaption the key action is configurable. The dedicated Magnification Control key provides fast toggling between two microscope magnification levels, low magnification for high level inspection of the fiber endface, and high magnification for detailed inspection of the fiber endface.



Fig. 15 Patch cord microscope components

- | | |
|---|---------------------------|
| 1 | FMAE adapter |
| 2 | QuickCapture™ key |
| 3 | Focus Control |
| 4 | Magnification Control key |

FMAE series adapters for the PCM

SmartClass™ Fiber devices with the PCM use FMAE series adapters to ensure consistent and accurate inspection for a wide variety connectors and applications. All PCM configurations ship with an included 2.5 mm interface. Kitted configurations may include additional FMAE adapters.

Features available with the P5000i Digital Probe

The P5000i Digital Probe is a portable handheld microscope used to view and inspect both the bulkhead (female) and patch cord (male) sides of fiber connectors, as well as other optical devices, such as transceivers.

The Digital Probe is specially designed to fit and operate comfortably and easily in-hand, allowing the user to inspect hard-to-reach connectors that are installed on the back side of patch panels or inside hardware devices. In order to support a wide variety of fiber optic connectors the P5000i provides an exchangeable FBPT inspection tip. The dedicated QuickCapture key provides either instant triggering of a pass/fail test or of freezing the live image. For best workflow adaption the key action is configurable. The dedicated Magnification Control key

provides fast toggling between two microscope magnification levels - low magnification for high level inspection of the fiber endface and high magnification for detailed inspection of the fiber endface.

The P5000i Digital Probe kit sold with the OLS-85 contains the standard barrel assembly (FBPP-BAP1), standard patch cord tips, and standard bulkhead tips.



Fig. 16 P5000i Digital Probe components

1	Inspection tip
2	Barrel assembly
3	QuickCapture™ key
4	Focus Control
5	Magnification Control key
6	USB 2.0 connector
7	1.83 m (6 ft) coil cable

FBPT series tips for the P5000i

The P5000i Digital Probe uses FBPT series inspection tips to ensure consistent and accurate inspection for a wide variety of connectors and applications. These connector-specific and universal inspection tips are interchangeable, which allows the P5000i Digital Probe to interface with different types of fiber connectors.




Fig. 17 FBPT series tips for the P5000i

1	Inspection tip
2	Barrel assembly

P5000i connection

The Probe application requires a P5000i Digital Probe in order to be fully functional (see the list of all accessories in [“Digital Probe” on page 67](#)).

1. Plug your P5000i into an USB port.
2. Connect the P5000i with the fiber being inspected.
3. Press the **[↶]** key, then tap the **[Probe]** button or use the arrows keys.
– or –
Press the **[≡]** key, to toggle between the measurement view and the inspection view.
The P5000i will display the live view.

QuickCapture™ key

In order to support different workflows the functionality of the QuickCapture™ key is configurable. Pressing the key will either freeze the live image or start a test:

Test	Pressing the key will automatically freeze the live image and perform a test.
Freeze	Pressing the key will automatically freeze the live image.

Camera mode

Set the QuickCapture™ key on the Digital Probe to get a live picture of the result.

NOTE: Use the Focus Control key to adjust the focus of the image.

✓ The Digital Probe is connected to the instrument.

- To display the live view, tap the [Probe] button.

File toolbar

Saving a picture

It is possible to save the frozen picture from the Digital Probe.

1. Press the [F] key.
2. Edit the file name.
3. Tap the [OK] button.
File is saved in the preselected group.

Configuring the Digital Probe

1. Tap the [Probe] button.
2. Press the [≡] key to configure the Digital Probe according to the following description.

Brightness settings

1. Set the luminosity manually using the arrow keys.
– or –
Tap the key [≡] > [Brightness] to tune the luminosity to the desired value.
2. Tap the [OK] button.
– or –
To confirm, press the center key within the arrow keys.


Test mode


Launching a test of the connector and fiber end-face

See “Freeze mode” on page 52 for more information on commands available.

To launch the test with the predefined profile (see “Profile” on page 52):

1. With the P5000i Digital Probe, use the Focus Control key (see Fig. 16) to adjust the image quality and sharpness.


2. Press the  key and tap the **Test** button.



– or –

Press the QuickCapture™ key on the Digital Probe if the button action is set to **Test**.

When the test procedure terminates, the information shown on the display depends on the current overlay setting:


- Press the  key, and then the **Overlay** button to change the overlay view. Repeat the action until the desired view appears.


Three views can be displayed: without overlay, with colored edges, and colored edges with a legend.

Overlay

A summary of test results is displayed at the upper right of the screen.

- **Zone A:** Represents the core zone. It is the area surrounding the core
- **Zone B:** Represents the cladding zone. It surrounds the majority of the fiber cladding.
- **Zone C:** Represents the epoxy ring.
- **Zone D:** Represents the ferrule zone. It identifies a portion of the ferrule near and around the fiber.



NOTE: To return to a Live image, press the QuickCapture™ key or press the  key and tap the **Live** button to view both the live image and a test result simultaneously.

If the test has been passed, the image has a green colored frame. If failed, the frame is red.




Fig. 18 Pass/Fail analysis – overlay view

Magnification Control

The Magnification Control allows you to modify the live display from high to low magnification and vice-versa. In the high magnification mode an automatic centering is available.

Picture selected: Live

Test Allows you to launch a (new) test of the connector (see [“Launching a test of the connector and fiber end-face” on page 49](#))

Freeze Allows you to freeze the live image and to save it later on the disk. Once this key is pressed, the freeze image is automatically set to position 2.

Picture selected: Test

Save allows you to save the selected picture in the current selected group in the disk of the instrument. Press the **[F10]** key, enter a name for the JPG file, and validate.

Overlay allows you to display or hide the limits of each zone and the defaults on image from a test result.

Magnification Control allows you to switch all the images from Digital Probe test results from high to low magnification and vice-versa.

Saving the test result in a JPG file

Once the test has been performed and the result is displayed on the OLS-85 screen:

1. Click on the **[F10]** key to save a JPG file of the test result in the currently active group on the OLS-85.
2. On the edition keypad, enter the name of the JPG file.
3. Press Enter to validate and save the document.

The file is saved in the current active group directory on the OLS-85.

Freeze mode

Once the image is acceptable (sharpness, luminosity are tuned correctly), you may freeze the picture. This feature allows you to store the resulting picture in order to compare it later to others or save it in a file. In Freeze Mode the picture has a blue colored frame.

NOTE: Freezing a result does not store the picture in a file (see [“Saving a picture” on page 49](#)). The result will be lost if the instrument is shut off.

Depending on the QuickCapture™ key setting, this key allows you to freeze the picture or to start a test.

Profile



1. Press the [≡] key.
2. Tap the [Profile] button.
3. Tap the Profile which will be used for the test of the fiber connector:
 - **E2000:** Pass/Fail criteria for precision metal ferrule connectors.
 - **Ribbon_:** Pass/Fail criteria for single-mode and multi-mode connectors from IEC 61300-3-35 standard.
 - **SM-UPC:** Pass/Fail criteria for single-mode UPC connectors from IEC 61300-3-35 standard.
 - **SM-APC:** Pass/Fail criteria for single-mode APC connectors from IEC 61300-3-35 standard.
 - **SM_PC:** Pass/Fail criteria for single-mode PC connectors from IEC 61300-3-35 standard.
 - **MM_:** Pass/Fail criteria for multi-mode connectors from IEC 61300-3-35 standard.
 - **SFP_:** Pass/Fail criteria for small form-factor pluggables.

Profiles contain the analysis parameters from which Pass/Fail criteria are determined.


Tip

1. Press the [≡] key.
2. Tap the [Tip] button.
3. Select the tip set on the Digital Probe to connect fiber for inspection and tap on the tip set.

More

1. Press the  key.
2. Tap the  button to view or change the storage location, settings and information of the Digital Probe.

Saving Probe/PCM results

Images can easily be saved by pressing the  key. Each time the key is pressed, the results will be stored.

NOTE: Instruments with an integrated patch cord microscope ("PCM") use a specific app, also called "PCM", to operate that microscope. Therefore, images taken by the PCM can easily be distinguished from images taken by an external digital probe.

NOTE: If the instrument with integrated patch cord microscope is connected to a Digital Probe, the images for the PCM and the Digital Probe will be stored in different folders.


Background

Every saving covers the complete image and overlay data available, regardless of the current screen contents:


- Two JPEG files for high and low magnification
- One XML file for test result description
- One PGM file (compressed) for failed locations
- One container file

When data is requested by the user (e.g. using SCPI), the provided jpg-file is created just in time by the software. This concept helps to save memory space and allows to maintain sharp zone rings for all resolutions (VGA, QVGA and QQVGA).

NOTE: Saved image and overlay data (JPEG, XML, PGM files) can be transferred, displayed and managed via the Smart Reporter software.

1. Select the desired overlay mode.
2. Press the  key.
In live view this action triggers the snapshot.

As file name the label prefix defined in the project settings is showed on top of the display.



3. To edit the file name, tap the name, edit it and tap [OK].
4. To change the ID, tap [Decrement ID] or [Increment ID].
5. Press [Save] to save the results.

The file is saved in the current active group directory.

8 DATA MANAGEMENT

NOTE: Results are always stored under the currently selected (active) project. Thus, to display stored results the desired project must be set to active first (see “Managing projects” on page 36).

Saving results

Saving results is explained in the descriptions of each application.

Data management of Probe and PCM tests

Recalling stored test results

Stored test results are displayed directly from the menu in the **Probe/PCM** application.

✓ The **Probe** or **PCM** application is selected.

1. Press the [≡] key.
 2. Tap the [More] button.
 3. Tap the [Measurement Data] button.
- The list of stored test results is displayed.*

Fiber ID	Pass/Fail	MM (IE)
FiberID1	--	MM (IE)
FiberID2	FAIL	MM (IE)
✓ Rem7	PASS	MM (IE)
Rem7	--	MM (IE)

To show all columns of the overview:

- Press the right/left arrow keys to show additional information. The displayed information depends on the selected application.

Fiber ID	Test label name
Pass/Fail	Passed/failed label


Profile	Profile label
Adapter	Type of adapter
Timestamp	Time and date of the test

Selecting/Deselecting test results

Test results first must be selected before they can be viewed or deleted. A selected entry is flagged by the check mark in the first column.

To select one test:

1. To select a test, tap it twice.
– or –
Press the arrow keys to highlight the test and press the center key
The check mark shows the selected test.



To select multiple tests:

- Repeat step 1 to select multiple entries.

To select all tests:

- Press the [≡] key and tap [Select All].

To deselect all tests:


- Press the [≡] key and tap [Deselect All].

Displaying test results


NOTE: When selecting multiple tests, all selected tests can be viewed one after another without switching back to the list and selecting another test.

To display a test result:


1. Select one, multiple or all tests.

2. Press the  key and tap the [View Selected] button.
The test results are displayed in full screen.


To view the next test result:

- ✓ The selected test result is displayed in full screen.
- 1. Press the center key.
– or –
- 2. Press the  key and tap the [Next] button.


To view the previous test result:

- ✓ The selected test result is displayed in full screen.
- ▶ Press the  key and tap the [Prev] button.


Sorting test results

- ✓ The measurement data overview is displayed.
- 1. Press the  key.
- 2. Tap the [More] button, then tap the [Sort Order] button.
- 3. Select the desired sorting order.


Show/hiding the overview columns

- ✓ The measurement data overview is displayed.
- 1. Press the  key.
- 2. Tap the [More] button, then tap the [Shown Columns] button.
- 3. Tap any column header to toggle its hide/show status.

Deleting stored results from a project

- ✓ The measurement data overview is displayed.
- 1. Select one, multiple or all tests.
- 2. Press the  key.
- 3. Tap the [Delete Selected] button.
The selected test results are deleted.

Actions when viewing images



To toggle the image magnification low/high

- ▶ Press the [≡] key, then tap the [Mag.] button.
 - or –
 - Tap the image.

To move the image to the area of interest:

- ▶ Sweep the high magnification image with your finger.
 - or –
 - Press the arrow keys to move the display to the desired direction.


Exporting results to USB

The export function allows to store CSV files and files used exclusively by the Smart Reporter on a USB stick.

To export files:

1. Connect a USB stick to one of the USB ports on the right side of the device.

The  symbol is displayed at the top of the screen.

2. Tap the  button.
3. Tap the [Data Storage] button.
4. Tap the [Export Results to USB] button.

The files are stored in the following directory structure:

storage_<device_type><serial_number>

<current_date>T<current_time>Z.

*<current_date>T<current_time>Z_<current_date>_001.csv
default*

<Label>.<Test_No>.<Application>

The CSV file contains the results of a measurement/of various measurements. The application file can be processed by the Smart Reporter.

Making a report

In order to make a report, please download the FiberChekPRO™ or SmartReporter software from the Viavi web site <http://updatemyunit.net>.

- Connect your instrument to your PC via the USB port and follow the instructions on the screen.

9 MAINTENANCE



⚠ WARNING

Dangerous voltage and invisible laser radiation

Maintenance or cleaning of the instrument while it is connected or operating may damage the instrument or injure you.

- ▶ Make sure that the instrument is switched off and disconnected from all power sources and optical radiation sources before maintenance or cleaning.

Cleaning the test port

It is a good idea to check that the optical connections are clean and clean them if necessary before starting measurements. Even very small dust particles on the end surfaces of the plugs or in the test adapters can adversely affect the accuracy of the measurement.

1. Switch off the instrument.
2. Remove the test adapter from the optical connection.
The plug end surface is now accessible.
3. Wipe off the plug end surface using a cotton bud soaked in isopropanol.
This cleaning method is very effective and leaves no residue.
4. Blow out the test adapter with clean compressed air (available in spray cans, e.g. anti-dust spray).

NOTE: Cover the optical connections with the dust cap whenever they are not in use. This prevents them from getting dirty.

Cleaning the instrument

If the instrument gets dirty through use, you can clean it using a soft cloth moistened with a mild solution of detergent.

NOTICE

Water and cleaning fluids

The instrument may be damaged or destroyed if water or cleaning fluids penetrate it.

- ▶ Make sure that water or cleaning fluids do not penetrate the instrument.

10 REMOTE CONTROL

Remote Command Documentation

- ▶ Please visit the Viavi web site at <http://updatemyunit.net> for the latest Remote Command Documentation "SCF RC Docs.exe" (self extracting zip-file).

11 ENVIRONMENTAL COMPLIANCE

This equipment was tested and found to comply with EN61326-1, IEC 61000-4-2, IEC61000-4-3, IEC/CISPR11. The limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment.

This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications.

Operation of this equipment in a residential area is likely to cause harmful interference, in which case you will be required to correct the interference at your own expense.

The authority to operate this equipment is conditioned by the requirements that no modifications be made to the equipment unless the changes or modifications are expressly approved by Viavi.

NOTE: To comply with FCC RF exposure compliance requirements, a separation distance of at least 20 cm must be maintained between the antenna of this device and all persons.

NOTE: This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

Industry Canada (IC)

This device complies with Industry Canada's licence-exempt RSSs. Operation is subject to the following two conditions:

- 1) This device may not cause interference;
- 2) This device must accept any interference, including interference that may cause undesired operation of the device.

NOTE: Device operation in the band 5150-5250 MHz is for indoor use only.

EU Radio Equipment Directive

In accordance with Article 10.8 of the EU Radio Equipment Directive 2014/53/EU, the following table provides information on the frequency bands and the maximum RF transmit power of this product for sale in the EU:

Frequency range (MHz)	Channels used	Max. Transmit Power (dBm/mW)
2400-2483.5	1-13	ODFM: 19.9 dBm (97.7 mW) CCK: 17.9 dBm (61.7 mW)
5150-5250	36-48	22.9 dBm (195 mW)
5250-5350	52-64	22.9 dBm (195 mW) with TPC ¹⁾ 19.9 dBm (97.7 mW) with TPC
5470-5725	100-140	29.9 dBm (977 mW) with TPC 26.9 dBm (490 mW) non-TPC

1) TPC means Transmit Power Control

12 SPECIFICATIONS

OLS-85 specifications

Laser safety	IEC 60825-1:2007
Laser classification	CLASS 1 LASER PRODUCT
Power setting resolution	0.01 dB
Tracking error	± 1 dB

Wavelengths

BN 2313/01	1310/1550 nm ¹⁾ , PC
BN 2313/05	850/1300 nm MM & 1310/1550 nm SM ¹⁾ , PC
BN 2313/06	1310/1490/1550/1625 nm ¹⁾ , PC
BN 2313/22	1310/1550/1625 nm ¹⁾ , APC
BN 2313/26	1310/1490/1550/1625 nm ¹⁾ , APC

1) Typically ± 20 nm

Wavelength accuracy

1310 nm, 1550 nm	± 20 nm
1490 nm, 1625 nm	± 3 nm
850 nm, 1300 nm	± 20 nm

Spectral width

1310 nm, 1550 nm, 1490 nm, 1625 nm	< 5 nm
850 nm ¹⁾ , IEC 61280-4-1 Encircled-Flux compliant	50 nm typ.
1300 nm ¹⁾	< 170 nm

1) LED source

Fiber types

Single-mode	9/125
Multi-mode	50/125, 62.5/125

Output power range (typical)¹

Single port instruments

2 wavelengths instruments	
Laser (CW)	-3.00 to 0.00 dBm (± 1 dB)
Laser (Modulated, Auto λ)	-6.00 to -3.00 dBm (± 1 dB)
3 & 4 wavelengths instruments	
Laser (CW)	-6.00 to -3.00 dBm (± 1 dB)
Laser (Modulated, Auto λ)	-9.00 to -6.00 dBm (± 1 dB)

Dual port instruments

SM port	
Laser (CW)	-3.00 to 0.00 dBm (± 1 dB)
Laser (Modulated, Auto λ)	-6.00 to -3.00 dBm (± 1 dB)
MM port	
LED (CW)	-25.00 to -22.00 dBm
LED (Modulated, Auto λ)	-28.00 to -25.00 dBm

Signal stability²

Short term	± 0.02 dB (within 15 min.)
Long term	± 0.20 dB (within 8 h)

Modes³

Continuous wave (CW), Modulated (270 Hz, 1 kHz, 2 kHz), Auto λ .

¹ CW signal, T = 23 °C \pm 3 K, including connector, depending on quality of connector applied to the OLS

² At ambient temperature range -10 °C to +55 °C, $\Delta T = \pm 0.3$ K

³ Signal coding for automatic power meter wavelength detection. Works only with Viavi power meters e.g. OLP-85

General specifications

Fiber inspection	Via P5000i Digital Probe with auto Pass/Fail analysis
Live image	320 x 240 pixels, 8 bit gray, 10 fps
Display	High contrast 3.5" TFT color touchscreen
Display resolution	0.01 dB / 0.001 μ W
Measurement units	dB, dBm, W
Data memory	10000 measurement results
Data readout	Via client USB interface or Ethernet
Remote control capability	Via USB
Electrical interfaces	2 USB Type A, 1 USB Micro-B, Ethernet
Power supply	Four-way power: NiMH, dry batteries, RBP2 Li-Ion Battery Pack, PS4 Universal AC/DC Power Supply 12 V Internal charging for RBP2 Li-Ion Battery Pack
Optical connectors	Interchangeable: SC, FC, ST, LC, DIN Fixed: SC
Recommended recalibration interval	3 years
Size (H x W x D)	208 x 112 x 64 mm (8.2 x 4.4 x 2.5 in)
Weight	750 g
Operating temperature range	-5 °C to +45 °C
Storage temperature range	-25 °C to +55 °C

13 ORDERING INFORMATION

OLS-85/OLS-85P stand alone units

OLS-85 Optical Laser Source

1310/1550 nm, PC universal adapter	BN 2313/01
1310/1550/1625 nm, APC universal adapter	BN 2313/22
1310/1490/1550/1625 nm, PC universal adapter	BN 2313/06
1310/1490/1550/1625 nm, APC universal adapter	BN 2313/26
850/1300 nm MM, 1310/1550 nm SM, PC universal adapter	BN 2313/05

Included items

Stand-alone units

- SmartClass™ Fiber instrument
- Soft shoulder case for SmartClass™ Fiber + accessories
- Electronic tool kit with Operating Manual, datasheet and Smart-Reporter SW on USB flash drive
- Two optical adapters: SC type, or selectable SC, FC, DIN, ST, LC in universal version
- Quick Start Manual and Safety Instructions
- Eight dry batteries AA size

Accessories

PS4 Universal AC/DC Power Supply for SmartClass™ Fiber	BN 2305/90.01
RBP2 Li-Ion Battery Pack for SmartClass™ Fiber	BN 2305/90.02
Calibration report	BN 2307/90.03
UC4 hands free carrier for SmartClass™ Fiber	BN 2128/01
UC4P hands free carrier for SmartClass™ Fiber with PCM	BN 2128/02
SCASE2 soft shoulder case for SmartClass™ Fiber	BN 2128/03
USB cable USB-A to Micro USB	K 807

Digital Probe

P5000i Digital Probe Microscope with FiberChekPRO software	FBP-P5000I
--	------------

INDEX

A

- AC line plug adapter 20
- Activating
 - project 38
- Adjusting
 - display brightness 30
- Auto-Lambda
 - Source 42
- Auto-off, setting 30

B

- Batteries
 - Danger 18
 - Recharging 19
 - Replacing 18
 - Tips 20
- Battery operation 11

C

- Cables, connecting 22
- Calibrating the touchscreen 32
- Changing system settings 29
- Cleaning
 - Instrument 60
 - Optical connections 60
 - Test port 60
- Common features 7
- Condensation 14
- Connecting optical cables 22
- Connector panel 16
- Control panel 24
- Creating
 - new project 36

D

- Damages during shipping 14
- Date & time, setting 31
- Deleting
 - project 38
- Device information, showing 30
- Device overview 15

- Differences between the devices 7
- Display
 - elements (Source) 40
- Display brightness, adjusting 30
- Display elements 25
- Display-off, setting 31

E

- Editing
 - project 37
- Environmental Management Program 70
- Environmental protection 20
- Ethernet protocol, setting 32

F

- Factory default 32
- Firmware, update 34

L

- Language, selecting 31
- Laser on/off
 - Source 43
- Laser safety 11
- Link Data Mode 40

M

- Menu navigation 28
- Menus 25
- Modulation (Source) 42

N

- Navigating in the menus 28

O

- On/Off 24
- Operation
 - Probe/PCM 39
- Operation from AC power 20
- Output port (Source) 41

Overview 15

P

Package contents 14

Packing material 14

Power level

Source 43

Power supply 17

Probe/PCM 62

Project

activating 38

creating 36

deleting 38

editing 37

Projects, working with 36

Proper usage 10

PS4 Universal AC/DC Power Supply 12

R

Recalling test results

PCM & Probe 55

Recovery 14

Recycling 20, 71

Remote commands

Error messages 62

Resetting to factory default 32

S

Saving test results 53

Selecting a language 31

Setting

auto-off 30

date & time 31

display-off 31

Ethernet protocol 32

Shipping damage 14

Showing device Information 30

Source

Auto-Lambda 42

display elements 40

laser on/off 43

modulation 42

output port 41

power level 43

wavelength 41

Specifications

General 66

Switching on/off 24

System settings, changing 29

T

Test adapter, mounting 22

Test results

recalling (PCM & Probe) 55

saving (PCM & Probe) 53

Touchscreen, calibrating 32

U

Updating the firmware 34

W

Wavelength (Source) 41

Working with projects 36

PRODUCT REGULATORY COMPLIANCE

Viavi Environmental Management Program

Superb performance and high quality have always characterized Viavi datacom and telecom measurement technology products. In this same world-class tradition, Viavi has an established, proactive program of environmental management.

Environmental management is an integral part of Viavi's business philosophy and strategy requiring the development of long-term, productive solutions to problems in the key areas of economics, technology, and ecology.

A systematic environmental management program at Viavi is essential in regard to environmental policy and enhances cooperation between ourselves and our business partners.

The Viavi Environmental Management Program considers:

Product design and manufacture

Environmental restrictions and requirements are taken into account during planning and manufacture of Viavi products. This attention ranges from the raw materials and finished components selected for use and the manufacturing processes employed, through to the use of energy in the factory, and right on up to the final stages in the life of a product, including dismantling.

Hazardous materials

Viavi avoids or uses with care any hazardous or dangerous material in the manufacturing process or the end product. If the use of a dangerous material cannot be avoided, it is identified in product documentation and clearly labeled on the product itself.

Packaging materials

Preference is given to reusable or biodegradable single-substance packaging materials whenever possible.

Environmental management partnerships

Viavi encourages our customers and suppliers who take this responsibility seriously to join Viavi in establishing their own environmental management programs.

EU WEEE and Battery Directives

This product, and the batteries used to power the product, should not be disposed of as unsorted municipal waste and should be collected separately and disposed of according to your national regulations..



Viavi has established a take-back process in compliance with the EU Waste Electrical and Electronic Equipment (WEEE) Directive, 2012/19/EU, and the EU Battery Directive, 2006/66/EC.

Instructions for returning waste equipment and batteries to JDSU can be found in the WEEE section of **Viavi's Standards and Policies web page** (<https://www.viavisolutions.com/en-us/corporate/legal/policies-standards#sustain>).

If you have questions concerning disposal of your equipment or batteries, contact JDSU's WEEE Program Management team at WEEE.EMEA@ViaviSolutions.com.

EU REACH

Article 33 of EU REACH regulation (EC) No 1907/2006 requires article suppliers to provide information if a listed Substances of Very High Concern (SVHC) is present in an article above a certain threshold.

For information on the presence of REACH SVHCs in Viavi products, see the Hazardous Substance Control section of **Viavi's Standards and Policies web page**.

EU CE Marking Directives (LV, EMC, RoHS, RE)

This product conforms with all applicable CE marking directives. Please see EU Declaration of Conformity for details.

California Proposition 65

California Proposition 65, officially known as the Safe Drinking Water and Toxic Enforcement Act of 1986, was enacted in November 1986 with the aim of protecting individuals in the state of California and the state's drinking water and environment from excessive exposure to chemicals known to the state to cause cancer, birth defects or other reproductive harm.

For the Viavi position statement on the use of Proposition 65 chemicals in Viavi products, see the Hazardous Substance Control section of **Viavi's Standards and Policies web page**.

"中国RoHS"

《电子信息产品污染控制管理办法》（信息产业部，第39号）
附录

本附录按照"中国RoHS"的要求说明了有关电子信息产品环保使用期限的情况，并列出了产品中含有的有毒、有害物质的种类和所在部件。本附录适用于产品主体和所有配件。

环保使用期限：



本标识标注于产品主体之上，表明该产品或其配件含有有毒、有害物质（详情见下表）。
其中的数字代表在正常操作条件下至少在产品生产日期之后数年内该产品或其配件内含有的有毒、有害物质不会变异或泄漏。该期限不适用于诸如电池等易耗品。
有关正常操作条件，请参见产品用户手册。
产品生产日期请参见产品的原始校准证书。

有毒、有害物质的类型和所在部件

元器件 (Component)	有毒、有害物质和元素					
	铅 (Pb)	汞 (Hg)	镉 (Cd)	六价铬 (CR ⁶⁺)	多溴联苯 (PBB)	多溴二苯醚 (PBDE)
产品主体 (Main Product)						
印刷电路板组件 (PCB Assemblies)	X	○	○	○	○	○
内部配线 (Internal wiring)	○	○	○	○	○	○
显示器 (Display)	○	○	○	○	○	○
键盘 (Keyboard)	○	○	○	○	○	○
塑料外壳零件 (Plastic case parts)	○	○	○	○	○	○
配件 (Accessories)	○	○	○	○	○	○

○：代表该部分中所有均质材料含有的该有毒、有害物质含量低于SJ/T11363-2006标准的限值。
X：代表该部分中所有均质材料含有的该有毒、有害物质含量高于SJ/T11363-2006标准的限值。



North America
Latin America
China
Germany

+1 844-468 4284
+1 954 688 5660
+86 21 6859 5260
+49 7121 86 0

Viavi product specifications and descriptions in this document are subject to change without notice.
© 2019